On the Selection of Sampling Points for Myocardial T₁ Mapping

Mehmet Akçakaya,¹* Sebastian Weingärtner,^{1,2} Sébastien Roujol,¹ and Reza Nezafat¹

Purpose: To provide a method for the optimal selection of sampling points for myocardial T_1 mapping, and to evaluate how this selection affects the precision.

Theory: The Cramér–Rao lower bound on the variance of the unbiased estimator was derived for the sampling of the longitudinal magnetization curve, as a function of T_1 , signal-to-noise ratio, and noise mean. The bound was then minimized numerically over a search space of possible sampling points to find the optimal selection of sampling points.

Methods: Numerical simulations were carried out for a saturation recovery-based T_1 mapping sequence, comparing the proposed point selection method to a uniform distribution of sampling points along the recovery curve for various T_1 ranges of interest, as well as number of sampling points. Phantom imaging was performed to replicate the scenarios in numerical simulations. In vivo imaging for myocardial T_1 mapping was also performed in healthy subjects.

Results: Numerical simulations show that the precision can be improved by 13–25% by selecting the sampling points according to the target T_1 values of interest. Results of the phantom imaging were not significantly different than the theoretical predictions for different sampling strategies, signal-to-noise ratio and number of sampling points. In vivo imaging showed precision can be improved in myocardial T_1 mapping using the proposed point selection method as predicted by theory.

Conclusion: The framework presented can be used to select the sampling points to improve the precision without penalties on accuracy or scan time. Magn Reson Med 000:000–000, 2014. © 2014 Wiley Periodicals, Inc.

Key words: Cramér–Rao bound; precision; sampling points; parameter estimation; Fisher information; myocardial T_1 mapping; tissue characterization

INTRODUCTION

Quantitative mapping of various magnetic resonance (MR) parameters, such as longitudinal (T_1) or transverse

 (T_2) relaxation times may provide diagnostic information beyond conventional contrast-weighted images (1). These quantitative techniques have found a range of applications in neurological (2,3), oncologic (4,5), and cardiac (6,7) applications. In particular, recently, myocardial T_1 mapping (8–12) has been shown to provide a technique for assessment of interstitial diffuse fibrosis (13). Furthermore, if myocardial T_1 maps are acquired both prior to and following contrast injection, they can further be used to measure the extracellular volume fraction (14), which has shown utility for detection of diffuse myocardial fibrosis (6).

Quantitative mapping approaches rely on acquiring multiple images of the same anatomy with different contrast weightings, achieved by varying one or more imaging parameter(s). These parameters may include flip angles (15), echo times (16), and/or inversion/saturation times (17). Then, a closed-form expression, which characterizes how the magnetization evolves, as a function of the unknown parameters of interest and the known parameter(s) that is being varied, is voxel-wise fitted to these series of images. Clearly, the number of images in this series has to be greater than or equal to the number of unknown parameters of interest. Thus, multiple images are acquired in the series with different contrasts, which correspond to different sampling points on the magnetization curve that is parameterized by the unknown quantities.

To achieve a sufficient level of precision in the estimation, a high number of images are acquired with different choices of sampling points (18), typically with a uniform distribution along the curve (10,19), resulting in a long acquisition time. However, there is a degree of freedom on the selection of the location and the number of the sampling points on the parametric curve, which is controlled by the sequence designer. Furthermore, in myocardial parameter mapping, the respiratory motion has to be compensated for. In myocardial T₁ mapping, this is commonly achieved by performing breath-holds, typically ranging from 11 to 17 heartbeats (8,10,11,20). Thus, it is critical to improve the precision of the sequence with this constraint on the total scan time.

The precision of the T_1 estimation process can be characterized by the Cramér–Rao Bound (CRB), which provides a lower bound on the variance of an unbiased estimator of a deterministic parameter (21). Furthermore, it characterizes when an estimator that achieves this lower bound exists. CRB has been used in a number of MRI applications, including diffusion kurtosis imaging (22,23) and parameter mapping (24–26). By considering a

¹Department of Medicine, Beth Israel Deaconess Medical Center and Harvard Medical School, Boston, Massachusetts, USA.

²Department of Computer Assisted Clinical Medicine, Computer Assisted Clinical Medicine, University Medical Center Mannheim, Heidelberg University, Mannheim, Germany.

Grant sponsor: NIH; Grant number: R01EB008743-01A2; Grant sponsor: NIH; Grant number: K99HL111410-01.

^{*}Correspondence to: Mehmet Akçakaya, Ph.D., Beth Israel Deaconess Medical Center, 330 Brookline Ave, Boston, MA 02215. E-mail: makcakay@bidmc.harvard.edu

Received 27 January 2014; revised 1 April 2014; accepted 16 April 2014 DOI 10.1002/mrm.25285

Published online 00 Month 2014 in Wiley Online Library (wileyonlinelibrary. com).

parametric model involving the gradient vector directions, as well as the diffusion and kurtosis tensors, an estimation theoretic approach was used to optimize the direction and *b*-value of the diffusion weighting gradients (23). It was shown that the optimal selection was different than the conventionally used set of standard gradient settings, and this selection lead to improvements in precision. In parameter mapping, an approach for determining experimental parameters were proposed in (24) for rapid simultaneous T_1 and T_2 measurement. This CRB-based technique was used in (25) for characterizing optimal sampling times to determine the T₂ relaxation time. The optimal scheme yielded a distribution concentrated at certain points related to the T_2 value, unlike the traditional uniform distribution utilized in other studies. Most similar to our current work is that of (26), which derived the CRB for T_1 mapping. This was then used to characterize the selection strategies for certain ordered sampling schemes, which lead to tractable analytical expressions. Because the search space in (26) was limited to these ordered sampling schemes, the resulting sampling schemes were not similar to that of (25) in T_2 mapping, which exhibited concentrations at certain areas of the parametric curve. Furthermore, only numerical simulations were used in (26), with no validation for phantom or in vivo data.

In this work, we sought to use an estimation theoretic framework for bounding the precision of a given myocardial T_1 mapping sequence in the presence of noise. We extend the bounding approach of (26) to accommodate a wide range of T_1 values of interest, by providing a Bayesian lower bound on the precision of the T_1 estimator, defined as the standard deviation of error for a given noise level. We then propose to determine the optimal selection of sampling points by numerically minimizing this bound. To test the theory in practice, we choose a saturation-based T_1 mapping sequence, and minimize the corresponding closed-form expression for the lower bound over a large search space of sampling points to select an improved distribution of sampling points for a given range of T_1 values of interest. Numerical experiments, phantom, and in vivo imaging are performed to evaluate the proposed sampling point strategy versus a uniform distribution of sampling points along the curve.

THEORY

Cramér-Rao Lower Bound for Parameter Mapping

For completeness, we describe the signal model and derive the CRB for T_1 mapping, which is similar to the derivation of (26). In myocardial T_1 mapping, at each sampled image voxel, we observe

$$y_k = M(x_k; a, b, T_1) + n_k,$$
 [1]

for $k \in \{1, ..., K\}$ representing images with different contrast weightings along the longitudinal magnetization curve. n_k is measurement noise. Furthermore,

$$M(x_k; a, b, T_1) = a(1 - b \exp(-x_k/T_1)), \quad [2]$$

is the magnetization at x_k on the magnetization curve, parameterized by a, the signal-to-noise ratio (SNR) of the image with fully-recovered magnetization (with appropriate normalization for the noise variable n_k); b, which captures the saturation/inversion efficiency, the effects of imaging pulses on the magnetization recovery (10) and the mean of noise in the magnitude images (27), and T_1 , the longitudinal relaxation time.

For the *K* observations, we let $\mathbf{Y} = \{y_1, y_2, ..., y_k\}$. Based on the noise distribution, the probability density function $p(\mathbf{Y} | a, b, T_1)$ can be determined. This in turn can be used to find the Fisher information matrix (21), whose info (i, j)th entry is given by:

$$[\mathbf{I}]_{ij} = -E \left[\frac{\partial^2}{\partial \theta_i \partial \theta_j} \log p(\mathbf{Y}|\mathbf{\theta}) \right],$$
[3]

where for ease of notation, $\theta_1 = a$, $\theta_2 = b$, $\theta_3 = T_1$. Then, the CRB states, if $p(\mathbf{Y} | a, b, T_1)$ satisfies certain regularity conditions (21), the estimator of T_1 , denoted by T_1^{est} , has variance bounded by

$$\operatorname{var}\left(T_{1}^{\operatorname{est}}\right) \geq \left[\mathbf{I}^{-1}\right]_{3,3}.$$
[4]

As the Rician noise in the magnitude images is wellapproximated as Gaussian noise with sufficient baseline SNR (27), and as the maximum-likelihood estimate for Gaussian noise corresponds to the least-squares estimator that is commonly used in the context of MR parameter estimation (2,8), we restrict ourselves to a Gaussian noise model. Without loss of generality, we consider unitvariance zero-mean Gaussian noise, as the non-zero mean can be captured using the parameters, *a* and *b* in the T₁ mapping model. In this setting,

$$\log p(\mathbf{Y}|\boldsymbol{\theta}) = C - \frac{1}{2} \sum_{k=1}^{K} |y_k - M(x_k; \boldsymbol{\theta})|^2, \quad [5]$$

where C is a constant. We also note that the maximumlikelihood estimator is unbiased and attains this lower bound for this log-likelihood function (21).

Optimization of Sampling Point Selection

The model described by Eqs. [1] and [5] leads to:

$$\mathbf{I}_{ij} = -E\left[\frac{\partial^2}{\partial\theta_i \partial\theta_j} \log p(\mathbf{Y}|\boldsymbol{\theta})\right] = \sum_{k=1}^{K} \frac{\partial M(x_k;\boldsymbol{\theta})}{\partial\theta_i} \frac{\partial M(x_k;\boldsymbol{\theta})}{\partial\theta_j}, \quad [6]$$

Thus, the Fisher information matrix for the T_1 mapping model of Eq. [6] is

$$\mathbf{I} = \begin{bmatrix} \sum_{k=1}^{K} \left(1 - be^{\frac{-x_k}{T_1}} \right)^2 & \sum_{k=1}^{K} \left(-ae^{\frac{-x_k}{T_1}} \right) \left(1 - be^{\frac{-x_k}{T_1}} \right) & \sum_{k=1}^{K} -ab\frac{x_k}{T_1^2} e^{\frac{-x_k}{T_1}} \left(1 - be^{\frac{-x_k}{T_1}} \right) \\ \sum_{k=1}^{K} \left(-ae^{\frac{-x_k}{T_1}} \right) \left(1 - be^{\frac{-x_k}{T_1}} \right) & \sum_{k=1}^{K} \left(-ae^{\frac{-x_k}{T_1}} \right)^2 & \sum_{k=1}^{K} a^2 b\frac{x_k}{T_1^2} e^{\frac{-2x_k}{T_1}} \\ \sum_{k=1}^{K} -ab\frac{x_k}{T_1^2} e^{\frac{-x_k}{T_1}} \left(1 - be^{\frac{-x_k}{T_1}} \right) & \sum_{k=1}^{K} a^2 b\frac{x_k}{T_1^2} e^{\frac{-2x_k}{T_1}} & \sum_{k=1}^{K} \left(abe^{\frac{-x_k}{T_1}} \frac{x_k}{T_1^2} \right)^2 \end{bmatrix},$$
[7]

The goal of the optimization procedure is to minimize the variance of the estimator of T_1 . From Eq. [4], var $(T_1^{\text{est}}) \ge [\mathbf{I}^{-1}]_{3,3}$, where

$$J(a, b, T_1, \{x_k\}) = [\mathbf{I}^{-1}]_{3,3}$$

= $\frac{\mathbf{I}_{11}\mathbf{I}_{22} - \mathbf{I}_{12}^2}{\mathbf{I}_{11}(\mathbf{I}_{22}\mathbf{I}_{33} - \mathbf{I}_{23}^2) - \mathbf{I}_{12}(\mathbf{I}_{33}\mathbf{I}_{12} - \mathbf{I}_{23}\mathbf{I}_{13}) + \mathbf{I}_{13}(\mathbf{I}_{12}\mathbf{I}_{23} - \mathbf{I}_{22}\mathbf{I}_{13})}.$
[8]

We minimize $J(a, b, T_1, \{x_k\})$ as a surrogate for minimizing var (T_1^{est}) . While this is a closed-form expression in terms of $\{x_k\}$ given the values of $[a, b, T_1]$ that we are interested in, the analytical minimization of the objective function is not tractable. Thus, we propose to solve

$$\left\{x_k^{\text{samp}}\right\} = \arg\min_{\left\{x_k\right\}} J(a, b, T_1, \left\{x_k\right\})$$
[9]

numerically over a grid of values for $\{x_k\}$ for the desired values of a, b, T_1 . Furthermore, in Appendix A, we show that

$$J(a, b, T_1, \{x_k\}) = \frac{1}{(ab)^2} J(1, 1, T_1, \{x_k\}).$$
 [10]

Thus the values of a, b only scales the lower bound itself, and the minimization over the grid of values for $\{x_k\}$ is independent of a, b, and only depends on the T_1 value of interest.

METHODS

All imaging was performed on a 1.5T Philips Achieva (Philips Healthcare, Best, The Netherlands) system. Phantom imaging was performed with a body coil, and in vivo imaging was performed with a 32-channel cardiac phased-array receiver coil. For this HIPAAcompliant study, the imaging protocol was approved by our institutional review board, and written informed consent was obtained from all participants.

Saturation and inversion recovery are the most commonly used techniques for sampling of the longitudinal magnetization recovery. Currently, there are no myocardial T_1 mapping sequences that sample an unperturbed inversion recovery curve in a single breath-hold, although there are Look-Locker-based variants (8,9). Hence, we concentrate on the saturation recovery (SR) curve. We choose a SR-based- T_1 mapping sequence, SASHA (10), for validation in this study. This sequence is electrocardiogramtriggered, and applies a saturation pulse at every heart beat, followed by imaging after a recovery time of x_k . The first heart beat is acquired without magnetization preparation, corresponding to an infinite SR time. As different SR times can be selected for each heart beat in this sequence, this increases the degrees of freedom for the selection of the optimal sampling points.

Numerical Simulations

Simulation Setup

We note from Eq. [8], $J(a, b, T_1, \{x_k\})$ depends directly on $1/a^2$, that is, the lower bound on the variance of the estimator scales inversely with the square of the SNR. Thus, the choice of *a* does not affect the evaluation of the function, only changing the scaling. Hence, a = 40 was arbitrarily chosen as the baseline SNR for the simulations. b = 0.9 was used for the simulations, selected from the typical experimental range of *b* values for SASHA between 0.9 and 1.1 (10). A trigger delay of 780 ms at 60 bpm and an acquisition window of 190 ms were used. Including the duration of the saturation pulse and the duration of the read-outs to the k-space center (with the assumption of a linear profile order), the allowable saturation times ranged between $T_{\min} = 140$ ms and $T_{\max} = 760$ ms.

Experiments

Two sets of experiments were performed. In Numerical Experiment A, the bound was evaluated for various T_1 values of interest for K=11 sampling points: (i) 1250 ms (myocardium precontrast), (ii) 450 ms (myocardium post-contrast), (iii) 950–1250 ms (precontrast T_1 range), (iv) 400–600 ms (postcontrast T_1 range), and (v) 450 and 1250 ms (myocardium precontrast and postcontrast). In Numerical Experiment B, the bound was evaluated for various K values of {5, 7, 9, 11, 13, 15} for T_1 values of interest from 950 to 1250 ms.

Additional experiments, the effects of changing T_{max} , and allowing for multiple sampling of the point at infinity were performed, and are included in Appendix B.

Numerical Optimization

The *K* sampling points $\{x_k\}$ were selected numerically as follows: The search space was generated as all the possible saturation times from T_{\min} to T_{\max} with a step size of 10 ms, and a saturation time at infinity, corresponding to

sampling without any magnetization preparation. Every possible ordered samplings of length K, allowing repetitions, was generated from the search space. The point at infinity was only allowed to be included once in any sampling as any imaging pulse performed during acquisition disturbs the magnetization at this point. Thus, to sample the same point at infinity requires multiple rest periods for the longitudinal magnetization to regrow, which decreases efficiency, especially considering the constraints on scan duration with the breath-hold acquisition. For every ordered sampling, $J(a, b, T_1, \{x_k\})$ was evaluated for the given $[a, b, T_1]$.

For the T_1 values in a continuous range, a Bayesian CRB variant was used (28), where

$$J_B(\theta_1, \theta_2, \{x_k\}) = \frac{1}{T_1^{\max} - T_1^{\min}} \int_{T_1^{\min}}^{T_1^{\max}} J(\theta_1, \theta_2, \theta_3, \{x_k\}) d\theta_3$$
[11]

was evaluated instead. The integral was performed numerically with a step size of 5 ms, while making sure the (nearly)-uniform distribution on T_1 was 0 at the boundary points as required in (28). For the bimodal T_1 value of optimization (v), the procedure was evaluated with both equal weights and weights proportional to $1/T_1^2$ for the variances associated with the two T_1 values.

The CRB on precision, defined as the square root of $J(a, b, T_1, \{x_k\})$ and denoted by CRB_{prec} , was evaluated for the points selected by the optimization procedure, and compared to a uniform distribution of sampling points, where K - 1 sampling points are uniformly spread in the range T_{\min} to T_{\max} , and a point at infinity as in (10).

Phantom Imaging

Imaging Setup

To characterize the effect of the choice of sampling points on the precision of the T_1 estimates, phantom imaging was performed using 14 NiCl₂ doped agarose vials (29), whose T_1 and T_2 values spanned the ranges of values found in the blood and myocardium precontrast and postcontrast. A single-shot steady-state free precession sequence with the following parameters was used: 2D single-slice, fieldof-view (FOV) = 210 × 170 mm², in-plane resolution = 1.9 × 2.5 mm², slice-thickness = 8 mm, TR/TE = 2.7/1.35 ms, flip angle = 70°, 10 ramp-up pulses, acquisition window = 190 ms, linear k-space ordering. All scans were repeated five times to average out random variations.

Experiments

The first set of experiments compared different sets of sampling points for K=11. The sets of sampling points tested were the ones chosen from the numerical simulations for T_1 values of interest varying from 950 to 1250 ms, from 400 to 600 ms, and 450 and 1250 ms. For comparison, acquisitions with saturation times uniformly distributed in the range T_{\min} to T_{\max} , plus a point at infinity (referred to as "uniform") were performed. Each scan was acquired with number of signal averages (NSA) = 5 for sufficient baseline SNR.

The second set of experiments evaluated the precision of the uniform and proposed point selection techniques for T_1 values of interest varying from 950 to 1250 ms, for different number of sampling points, $K = \{5, 7, 9, 11, 13, 15\}$. NSA = 5 was used for these scans as well.

The third set of experiments was performed to validate the inverse linear dependence of precision to baseline SNR. The sequence was designed to last for K=11 heartbeats, and the scans were acquired with NSA = {1, 3, 5, 10}. The uniform point selection and the proposed point selection for T_1 values of interest varying from 950 to 1250 ms were utilized.

In Vivo Imaging

Myocardial T_1 maps were acquired in five healthy adult subjects (four women, 23.4 ± 3.3 years) without a contrast injection. Images were acquired in the short-axis of the heart. The SASHA sequence was used with a singleshot acquisition and a single-shot steady-state free precession readout with the following parameters: 2D single-slice, $FOV = 300 \times 300 \text{ mm}^2$, in-plane resolution = $1.9 \times 2.5 \text{ mm}^2$, slice-thickness = 8 mm, TR/ TE = 3.1/1.55 ms, flip angle = 70°, 10 ramp-up pulses, SENSE rate = 2, acquisition window = 190 ms, linear kspace ordering. The sequence duration was 11 heartbeats, and it was performed during breath-hold. Acquisitions were performed using a uniformly distributed point selection and the proposed selection of points for T_1 values between 950 and 1250 ms. For each point selection strategy, the acquisition was repeated five times to average out random variations.

Image Analysis

 T_1 estimation was performed offline using MATLAB (v7.6, MathWorks, Natick, MA) using the 3-point model (10) of Eq. [2]. Any point acquired without a saturation preparation was assumed to be sampled at an infinite (∞) saturation time. Multiple points acquired at the same point on the magnetization recovery curve were used individually for fitting, instead of averaging prior to fitting.

For T_1 measurements, a region-of-interest (ROI) analysis was performed for both phantom and in vivo imaging. The mean value and standard deviation in the ROI was recorded for each acquisition. The estimated T_1 value, T_1^{est} , is reported as an average \pm standard deviation of the mean values in the ROI (over the five acquisitions for each sampling strategy), which is a surrogate for accuracy and the interscan reproducibility. Under the assumption that the T_1 values are homogenous in the small ROI (11), the precision, prec (T_1^{est}), is reported as the average \pm standard deviation of the spatial standard deviation of the T_1 values in the ROI (over five acquisitions for each sampling strategy).

For phantom imaging, circular ROIs were drawn across each of the 14 vials, starting from the center of the vial and containing ~300 voxels. The ratios of the mean prec (T_1^{est}) for the two sampling strategies were then compared to the ratios predicted by the theoretical model for different K and T_1 values in the union of the ranges of optimization using paired *t*-test. For the SNR experiment, the correlation between prec (T_1^{est}) for different NSAs was compared to $1/\sqrt{(NSA)}$ using Pearson's linear

Results of the Proposed	Optimization Procedure for Sampling Point Selection for Various T_1 Values of Interest and 11 Sampling Points
T 1 (

I ₁ values of	Sampling Point		
interest (ms)	Selection Strategy	Saturation Recovery Sampling Points (ms)	CRB _{prec} (ms)
1250	Proposed	(140, 140, 140, 140, 760, 760, 760, 760, 760, 760, ∞)	82.6
	Uniformly distributed	(140, 209, 278, 347, 416, 485, 554, 623, 692, 760, ∞)	111.3
450	Proposed	(140, 140, 140, 480, 480, 480, 480, 480, 480, 480, 4	33.9
	Uniformly distributed	(140, 209, 278, 347, 416, 485, 554, 623, 692, 760, ∞)	39.1
950-1250	Proposed	(140, 140, 140, 140, 760, 760, 760, 760, 760, 760, ∞)	71.8
	Uniformly distributed	(140, 209, 278, 347, 416, 485, 554, 623, 692, 760, ∞)	94.0
400-600	Proposed	(140, 140, 140, 510, 510, 510, 510, 510, 510, 510, ∞)	36.9
	Uniformly distributed	(140, 209, 278, 347, 416, 485, 554, 623, 692, 760, ∞)	42.0

The saturation recovery times are chosen between $T_{min} = 140$ ms and $T_{max} = 760$ ms (and possibly one point at ∞ , representing an acquisition with no preparation). The associated Cramér–Rao lower bound (CRB) on the standard deviation (std) of the T_1 estimator (CRB_{prec}) when using the proposed strategy and a strategy selecting uniformly separated points are also depicted. The results indicate that the optimization procedure favors selection of multiple points at the same locations, with multiple points at T_{min} , as well as multiple points close to the T_1 value of interest (or the closest to it if these values are greater than T_{max}). In all cases, the optimized procedure has a lower CRB associated with the variance and standard deviation of the T_1 estimator, when compared to a uniform distribution of sampling points.

correlation coefficient for both sampling strategies. For in vivo imaging, ROIs were manually drawn, by two independent experienced readers, in the left ventricle myocardium and the blood pool. T_1^{est} and $\text{prec}(T_1^{\text{est}})$ were calculated for both strategies.

RESULTS

Numerical Simulations

Point Selection for Different T_1 Values of Interest

The 11 sampling points selected by the proposed method for the first four different cases of T_1 values of interest, and the associated CRB_{prec} are depicted in Table 1 for SNR = 40, along with the corresponding CRB_{prec} for a uniform distribution of sampling points. The proposed choice of sampling points yields a trimodal distribution: 1 point at ∞ , 3–4 points at T_{\min} , and 6–7 points close to the T_1 value(s) of interest. The proposed selection leads to a reduction in CRB_{prec} in all cases (13, 26, 24, 12.1% reduction for T_1 values of interest of 450, 1250, 950 – 1250, 400 – 600 ms, respectively) compared to the uniform selection.

For the fifth case of T_1 values, the optimization with equal weights yielded 4 points at 140 ms, 6 at 760 ms, and 1 at $\infty.$ In this case, $\text{CRB}_{\rm prec}$ associated with the $T_1 = 450$ ms estimation was 42.3 ms, whereas CRB_{prec} for $T_1 = 1250$ ms was of 82.6 ms, where the former is worse than for uniform selection. Note that for equal weights, the variance of the higher T_1 value dominates the optimization procedure, yielding the same choice of points as for just $T_1 = 1250$ ms. When weights of $1/T_1^2$ are used to capture the percentage variation in estimation, the optimal selection becomes 4 points at 140 ms, 4 at 570 ms, 2 at 760 ms, and 1 at ∞ , resulting in reductions in CRB_{prec} of 5 and 18% (for T_1 of 450 and 1250 ms, respectively) compared to uniform selection. This latter selection of points is used in subsequent imaging for these T_1 values of interest.

Effect of Changing the Number of Sampling Points

Table 2 depicts the results of the proposed point selection strategy and CRB_{prec} for both point selection strategies for T_1 values of interest of 950–1250 ms for various

Table 2

The Variation of the Selected Points for Different Number of Sampling Points, K, as well as the Associated CRB on the Standard Deviation of the T_1 Estimator (CRB_{prec}) when Using the Proposed Strategy and a Strategy Selecting Uniformly Separated Points

# sampling	Point Selection		CRB _{prec}
points, <i>K</i>	Strategy	Saturation Recovery Sampling Points (ms)	(ms)
5	Proposed	(140, 140, 760, 760, ∞)	117.7
	Uniform	(140, 347, 554, 760, ∞)	144.8
7	Proposed	(140, 140, 760, 760, 760, 760, ∞)	98.6
	Uniform	(140, 264, 388, 512, 636, 760, ∞)	130.4
9	Proposed	(140, 140, 140, 760, 760, 760, 760, 760, ∞)	88.8
	Uniform	(140, 229, 318, 407, 496, 585, 674, 760, ∞)	119.4
11	Proposed	(140, 140, 140, 140, 760, 760, 760, 760, 760, 760, ∞)	82.6
	Uniform	(140, 209, 278, 347, 416, 485, 554, 623, 692, 760, ∞)	111.3
13	Proposed	(140, 140, 140, 140, 140, 760, 760, 760, 760, 760, 760, 760, ∞)	78.2
	Uniform	(140, 196, 252, 308, 364, 420, 476, 532, 588, 644, 700, 760, ∞)	105.2
15	Proposed	(140, 140, 140, 140, 140, 760, 760, 760, 760, 760, 760, 760, 76	74.8
	Uniform	(140, 188, 236, 284, 332, 380, 428, 476, 524, 572, 620, 668, 716, 760, ∞)	99.6

The range of T_1 values of interest was from 950 to 1250 ms, with $T_{min} = 140$ ms and $T_{max} = 760$ ms. The proposed selection procedure again favors a trimodal distribution of points among T_{min} and T_{max} , which results in an improvement over the CRB for variance and std of the T_1 estimator with respect to a uniform distribution of selection points for all *K*.

	Unifo Distribute	ormly ed Points		Points Opt for $T_1 = 950$ -	imized 1250 ms			Points Opt for $T_1 = 400$	iimized - 600 ms			Points Optimi $T_1 = 450 \& 1$	ized for 250 ms	
	${\cal T}_1^{ m est}$	Prec(T, est)	$T_1^{\rm est}$	$Prec(T, ^{est})$	Prec wrt.	Theory Prec wrt.	$\mathcal{T}_{1}^{\mathrm{est}}$	$Prec(T_1^{est})$	Prec wrt.	Theory Prec wrt.	$T_1^{\rm est}$	Prec(T, ^{est})	Prec wrt.	Theory Prec wrt.
Vial	(ms)	(sm)	(ms)	(su)	Uniform	uniform	(sm)	(ms)	Uniform	Uniform	(su)	(ms)	Uniform	Uniform
-	1457 ± 7.7	69.2 ± 6.3	1456 ± 7.4	48.1 ± 5.2	0.70	0.71	1432 ± 12.7	72.0 ± 5.2	1.04	1.01	1450 ± 26.5	48.8 ± 4.4	0.71	0.81
2	$\textbf{1144} \pm \textbf{14.5}$	55.7 ± 6.7	1130 ± 7.1	41.0 ± 2.5	0.74	0.76	1127 ± 6.8	54.1 ± 2.7	0.97	0.98	1125 ± 12.9	44.7 ± 2.3	0.80	0.82
ო	1151 ± 11.5	53.2 ± 3.8	1155 ± 8.6	43.1 ± 1.4	0.81	0.76	1146 ± 12.2	54.0 ± 6.7	1.02	0.98	1147 ± 20.3	45.3 ± 1.9	0.85	0.82
4	274 ± 5.0	15.3 ± 1.2	275 ± 6.8	23.2 ± 2.5	1.51	1.71	271 ± 7.7	14.7 ± 0.8	0.96	0.99	$\textbf{269} \pm \textbf{4.5}$	15.0 ± 0.9	0.98	1.17
2	429 ± 6.6	17.3 ± 1.1	431 ± 2.3	18.8 ± 1.5	1.09	1.11	426 ± 6.8	14.1 ± 1.0	0.81	0.87	426 ± 5.9	15.2 ± 1.0	0.88	0.96
9	729 ± 10.3	31.3 ± 1.2	724 ± 2.1	26.1 ± 3.3	0.83	0.86	724 ± 4.6	29.0 ± 2.6	0.93	0.91	727 ± 13.9	27.7 ± 1.6	0.88	0.87
7	595 ± 6.6	20.0 ± 1.6	599 ± 6.0	$\textbf{18.8} \pm \textbf{1.6}$	0.94	0.94	590 ± 11.3	17.8 ± 1.9	0.89	0.89	588 ± 11.8	$\textbf{16.6} \pm \textbf{1.2}$	0.83	0.89
ø	597 ± 5.0	$\textbf{25.9}\pm\textbf{0.9}$	597 ± 5.0	23.3 ± 3.0	06.0	0.94	592 ± 5.3	22.8 ± 1.3	0.88	0.89	594 ± 11.8	23.1 ± 0.3	0.89	0.89
റ	980 ± 11.2	34.6 ± 1.2	981 ± 10.4	25.2 ± 0.6	0.73	0.78	971 ± 8.5	29.0 ± 2.5	0.84	0.96	983 ± 13.5	24.2 ± 1.0	0.70	0.84
10	823 ± 13.3	29.8 ± 2.1	822 ± 7.7	24.2 ± 2.0	0.81	0.83	816 ± 8.3	29.0 ± 2.4	0.97	0.93	816 ± 15.0	22.1 ± 1.3	0.74	0.86
1	1148 ± 18.4	$\textbf{52.6}\pm\textbf{6.4}$	1144 ± 8.3	37.8 ± 1.5	0.72	0.76	1129 ± 15.5	45.9 ± 2.4	0.87	0.98	1151 ± 18.7	42.0 ± 2.8	0.80	0.82
12	341 ± 4.1	19.4 ± 1.7	342 ± 5.8	$\textbf{22.6} \pm \textbf{2.4}$	1.17	1.29	337 ± 7.0	16.3 ± 1.0	0.84	0.89	335 ± 8.7	17.3 ± 0.3	0.89	1.01
13	963 ± 13.8	$\textbf{49.8} \pm \textbf{3.4}$	962 ± 6.2	35.4 ± 2.2	0.71	0.79	956 ± 3.7	47.1 ± 5.4	0.95	0.95	964 ± 8.6	37.9 ± 4.1	0.76	0.83
14	1130 ± 10.6	55.8 ± 5.6	1137 ± 10.5	45.0 ± 3.9	0.81	0.76	1121 ± 4.5	52.9 ± 2.2	0.95	0.98	1139 ± 15.8	45.0 ± 1.7	0.81	0.82
The rat form." all <i>T</i> ₁ v experin	o of the stand The points fro alues, respect iental ratios (<i>F</i>	dard deviation in numerical tively. The the p=0.26, $P=0$	of the T_1 esti- experiment #3 coretical predi- 0.09, $P = 0.05$	imator for each #4, and #5 sh ction correspor for the point s€	proposed s now improve nding to the	ampling stra ment on the ratios of sta different <i>T</i> ₁ v	tegy and that c standard devi indard deviatio alues of intere	of the uniform iation of the u ins are also re st, respectivel	i sampling st iniform samp ∋ported for ∉ ly) for vials v	rategy is rep ling strategy each samplir vith <i>T</i> ₁ value	orted as "prec r for <i>T</i> ₁ values ng strategy, wh s in the union	ision (prec) w > 600 ms, T_1 ich are in goo of the ranges	ith respect values < 11 od agreeme of optimiza	to (wrt) uni- 50 ms, and ent with the tion.

Results of the Phantom Imaging Over All 14 vials with Different T₁ Values for Various Sampling Strategies, Where Each Acquisition was Repeated five Times and K=11

Table 3

Table 4

The Results of Phantom Imaging for Various Values of Number of Sampling Points, K, Depicting the Precision Values, the Ratio of the Standard Deviation of the Proposed Point Selection for T_1 Values Ranging From 950 to 1250 ms and that of the Uniformly Distributed Point Selection, as Well as the Ratio Predicted by Theory

Vial #	Evaluation Metric	K=5	K=7	K=9	K = 11	K = 13	K = 15
2	Uniform - <i>prec</i> (T1 ^{est}) (ms)	70	59	58	56	50	47
	Proposed - <i>prec</i> (T1 ^{est}) (ms)	53	45	44	41	39	35
	Std ratio proposed vs. uniform	0.76	0.77	0.73	0.73	0.80	0.76
	Theoretical ratio	0.83	0.77	0.76	0.76	0.76	0.77
3	Uniform - <i>prec</i> (T ₁ ^{est}) (ms)	62	56	55	53	50	47
	Proposed - <i>prec</i> (T1 ^{est}) (ms)	53	45	44	43	40	37
	Std ratio proposed vs. uniform	0.86	0.81	0.79	0.81	0.79	0.77
	Theoretical ratio	0.82	0.76	0.75	0.75	0.76	0.77
9	Uniform - <i>prec</i> (T ₁ ^{est}) (ms)	43	36	35	34	34	30
	Proposed - <i>prec</i> (T1 ^{est}) (ms)	35	28	28	25	24	25
	Std ratio proposed vs. uniform	0.81	0.78	0.81	0.73	0.70	0.85
	Theoretical ratio	0.85	0.79	0.78	0.79	0.79	0.80
11	Uniform - <i>prec</i> (T1 ^{est}) (ms)	59	54	52	53	45	45
	Proposed - <i>prec</i> (T1 ^{est}) (ms)	49	39	41	37	35	34
	Std ratio proposed vs. uniform	0.84	0.73	0.80	0.71	0.79	0.76
	Theoretical ratio	0.82	0.76	0.75	0.75	0.76	0.76
13	Uniform - <i>prec</i> (T ₁ ^{est}) (ms)	53	49	48	49	44	40
	Proposed - <i>prec</i> (T1 ^{est}) (ms)	45	41	37	35	34	33
	Std ratio proposed vs. uniform	0.85	0.84	0.78	0.71	0.78	0.82
	Theoretical ratio	0.85	0.79	0.79	0.79	0.80	0.80
14	Uniform - <i>prec</i> (T ₁ ^{est}) (ms)	63	62	58	56	50	48
	Proposed - <i>prec</i> (T1 ^{est}) (ms)	53	44	45	45	39	36
	Std ratio proposed vs. uniform	0.84	0.70	0.79	0.81	0.77	0.75
	Theoretical ratio	0.82	0.77	0.76	0.76	0.76	0.77

The ratios are in good agreement across all vials considered (P = 0.97).

numbers of sampling points, K. The trend for the optimization strategy is to select 1 point at ∞ , floor(K/2)-1 or floor(K/2) points at T_{\min} , and floor(K/2) or floor(K/2) + 1 points at T_{\max} (as $T_1 > T_{\max}$). The reduction in CRB_{prec} compared to uniform sampling is 18.7, 24.4, 25.6, 25.8, 25.7, 24.9% for K=5, 7, 9, 11, 13, 15 points, respectively.

Phantom Imaging

Precision for Different Sampling Point Selections

Table 3 depicts the results of the phantom measurements over all 14 vials with different T_1 values for various sampling strategies with NSA = 5 for each acquisition and K=11. The estimated T_1 values for different sampling strategies for each vial are in good agreement with theory, even though there are variations between the five acquisitions due to noise (reflected as the standard deviation in the T_1^{est}). The ratio of $\operatorname{prec}(T_1^{\text{est}})$ for each proposed sampling strategy and that of the uniform sampling strategy is reported as "precision (prec) with respect to (wrt) uniform." The points optimized for T_1 values from 950 to 1250 ms show consistent improvement (as a ratio < 1) in this range with respect to the uniform strategy. Similarly, the points optimized for T_1 values from 400 to 600 ms show consistent improvement for smaller T_1 values < 1150 ms. The points optimized for two T_1 values of 450 and 1250 ms show improvement in standard deviation of the estimator for all T_1 values. Theoretical prediction for the ratios of $\operatorname{prec}(T_1^{\text{est}})$ with

respect to that of the uniform strategy is also reported for all selection of points. These are not statistically different than the experimental ratios (P=0.26, P=0.09, P=0.05 for T_1 values of 950 - 1250 ms, 400 - 600 ms, 450 and 1250 ms, respectively) for vials with T_1 values in the union of the ranges of optimization.

Precision versus Number of Sampling Points

Table 4 shows phantom results evaluating how prec (T_1^{est}) varies with number of sampling points, K. The ratio of prec (T_1^{est}) between the proposed and uniform point selections are depicted for $K = \{5, 7, 9, 11, 13, 15\}$ for the vials, whose T_1 values are between 950 and 1250 ms. Corresponding ratios predicted by theory for these vials are also reported. The experimental and theoretical ratios are in good agreement across all vials considered (P = 0.97).

Precision versus Baseline SNR

Figure 1 depicts prec (T_1^{est}) with changing NSA (equivalently $\sqrt{\text{SNR}}$) for all vials with T_1 values between 950 and 1250 ms and for both point selection strategies. The best linear fits (y = ax) are also shown (dashed lines) as a function of changing NSA. As predicted by Eq. [10], these are in good agreement with the experimentally measured prec (T_1^{est}) in all cases, with Pearson correlation coefficients >0.99 for all vials and for both point selection strategies. The baseline SNR was 29.2 ± 2.3 for the case of full longitudinal recovery (i.e., with SR time of ∞) image with NSA = 1.



FIG. 1. Standard deviation of the T_1 estimator as a function of varying NSA (or equivalently SNR) for the proposed point selection for T_1 values ranging from 950 to 1250 ms and the uniformly distributed point selection. The experimental results are in good agreement with the best linear fit of the form y = ax (dashed lines) with Pearson coefficients >0.99 for all cases. [Color figure can be viewed in the online issue, which is available at wileyonlinelibrary.com.]

In Vivo Imaging

Table 5 depicts the results of in vivo T_1 measurements for the five healthy subjects using the proposed and uniform sampling strategies. The reduction in prec (T_1^{est}) as measured by reader 1 were 13.9 and 20.1% for the myocardium and blood, respectively, with the proposed approach compared to the uniform selection strategy. The corresponding reductions measured by reader 2 were 20.2 and 26.6% for the myocardium and blood, respectively. Overall, there was a 17.0 and 23.5% reduction in the prec (T_1^{est}) in the myocardium and blood, respectively, using the proposed approach compared to the uniform selection.

Figure 2 shows example myocardial T_1 maps from a healthy subject (subject #5), acquired using the proposed and uniform point selection strategies at a heart rate of 60 bpm. Homogeneity of the T_1 maps is visibly improved in the blood pools, where the T_1 value is higher compared to the myocardium. prec (T_1^{est}) in the myocardium averaged over the five acquisitions showed 22 and 20% improvement with the proposed selection as measured by readers 1 and 2, respectively. Similarly, prec (T_1^{est}) for the blood pool showed 22 and 31% improvement as measured by readers 1 and 2, respectively. In comparison, the theory predicts reductions of 24.6 and 30.3% for CRB_{prec} of the myocardium and blood, respectively.

DISCUSSION

In this study, we used an estimation theoretic framework for evaluating the precision of myocardial T_1 mapping techniques to devise a sampling point selection strategy. This was achieved by minimizing the CRB on the variance of the estimators over the search space of possible sampling points. Subsequently, we evaluated this point selection strategy in a saturation-recovery-based T_1 mapping sequence, and compared it to a uniform distribution of sampling points along the SR curve. The improvements in precision in phantom and in vivo imaging matched those from the numerical simulations derived from the theory.

The proposed optimized selection procedure yielded sampling points that are concentrated in three areas of the T_1 relaxation curve: one at a low saturation (or inversion) recovery time (T_{\min}) , one with a saturation time near (or due to sequence limitations, T_{\max} , the closest possible value to) the T_1 value of interest, and the third point at ∞ . The multiplicity of the points at these locations varies with the total number of points, as well as the sequence being simulated (e.g., taking multiple points at ∞ may not be feasible in cardiac MR). In all cases, these lead to an improved precision with respect to a uniform distribution of the points along the T_1 relaxation curve. Thus, from a theoretical perspective it is better to improve the SNR at specific well-chosen points

Table 5

The Results of In Vivo Imaging on Five Healthy Subje	ects Using the Proposed and	d Uniform Sampling Strategies	, where Each Acquisition
was Repeated Five Times, as Measured by Two Inde	pendent Readers.		

				Uniform point selection		Proposed po	int selection	
	Subject	Heart rate (bpm)	Anatomy	T ₁ ^{est} (ms)	Prec(T1 ^{est}) (ms)	T ₁ ^{est} (ms)	Prec(T ₁ ^{est}) (ms)	Std wrt. uniform
Reader #1	1	65	Myocardium	1163 ± 24.2	104.9 ± 15.6	1155 ± 33.3	103.7 ± 11.3	0.99
			Blood	1926 ± 17.2	162.5 ± 15.3	1897 ± 26.8	114.7 ± 10.2	0.71
	2	75	Myocardium	1350 ± 19.6	177.9 ± 11.5	1348 ± 30.0	103.7 ± 11.3	0.89
			Blood	1751 ± 46.4	187.6 ± 30.3	1766 ± 59.6	149.0 ± 5.8	0.79
	3	70	Myocardium	1242 ± 40.7	153.4 ± 21.1	1255 ± 15.6	124.5 ± 14.5	0.81
			Blood	1799 ± 9.5	189.0 ± 20.7	1811.2 ± 36.5	155.0 ± 12.0	0.82
	4	68	Myocardium	1200 ± 33.9	110.9 ± 11.5	1193 ± 13.6	93.7 ± 9.9	0.84
			Blood	1733 ± 44.2	140.8 ± 14.8	1754 ± 16.1	128.1 ± 17.8	0.91
	5	58	Myocardium	1135 ± 77.9	105.1 ± 18.9	1185 ± 16.5	81.8 ± 10.3	0.78
			Blood	1757 ± 52.1	150.2 ± 15.1	1751 ± 15.1	116.8 ± 9.6	0.78
Reader #2	1	65	Myocardium	1188 ± 15.6	115.7 ± 16.2	1194 ± 15.3	91.7 ± 11.0	0.79
	•		BIOOD	1925 ± 25.1	166.7 ± 16.5	1903 ± 19.2	111.8 ± 7.7	0.67
	2	75	Myocardium	1314 ± 25.4	$1/4.4 \pm 28.6$	1307 ± 21.5	150.1 ± 21.2	0.86
		70	Blood	$1//2 \pm 24.1$	205.0 ± 32.3	$1/55 \pm 20.8$	148.2 ± 7.2	0.72
	3	70	Myocardium	1204 ± 70.2	117.0 ± 16.7	1218 ± 37.8	83.0 ± 13.5	0.71
			Blood	$1/87 \pm 37.3$	$1/9.7 \pm 26.5$	1809 ± 35.4	138.1 ± 14.9	0.77
	4	68	Myocardium	1213 ± 49.5	107.3 ± 16.7	1207 ± 31.7	85.0 ± 5.3	0.79
			Blood	1755 ± 34.7	161.4 ± 16.7	1780 ± 19.2	131.4 ± 12.8	0.81
	5	58	Myocardium	1168 ± 54.3	95.3 ± 6.8	1187 ± 14.8	76.6 ± 5.0	0.80
			Blood	1772 ± 47.8	164.7 ± 18.9	1761 ± 20.8	114.2 ± 10.4	0.69

 T_1^{est} is reported as the mean \pm std of the spatial average T_1 values in the ROI across five scans, as a surrogate for accuracy and interscan reproducibility. prec (T_1^{est}) is reported as the mean \pm std of the spatial std of the T_1 values in the ROI across five scans, as a surrogate for the precision within the scan. Std wrt. uniform is the ratio of the mean values of prec (T_1^{est}) using the proposed and uniform point selection, characterizing the percentage gain in precision. prec (T_1^{est}) in the myocardium and blood was reduced by 17.0 and 23.5%, respectively, using the proposed approach.

along the magnetization recovery curve by sampling multiple times at these locations than to distribute the sampling points along the curve. We note that this choice of points tends to be different than another existing hypothesis that points on the curve with higher magnitude derivatives with respect to time are more influential on the final estimate (30). The CRB for T_1 mapping was also derived previously in (26). This was then used to select a new set of sampling points, however, the search space was only limited to ordered (linear, quadratic, cubic) sampling schemes due to the analytical tractability of these expressions. Thus, the trimodal distribution resulting from our numerical optimization scheme with a more extensive



FIG. 2. T_1 maps from a healthy subject (subject #5), acquired using the proposed and uniform point selection strategies. The T_1 map generated using the proposed point selection is more homogenous, as visible in the blood pool. prec (T_1^{est}) averaged over the five acquisitions were 105.1 ± 18.9 and 81.8 ± 10.3 ms for the myocardium as measured by reader 1, and 95.3 ± 6.8 and 76.6 ± 5.0 ms as measured by reader 2 for uniform and proposed selection strategies, respectively. prec (T_1^{est}) for the blood pool were 150.2 ± 15.1 and 116.8 ± 9.6 as measured by reader 1, and 164.7 ± 18.9 and 114.2 ± 10.4 ms as observed by reader 2 for uniform and proposed selection strategies, respectively.

search space, was not observed in (26). We note that the trimodal distribution of optimal points presented here is similar to the bimodal distribution observed in T_2 mapping where a 2-parameter model was utilized (25). Apart from the differences in search spaces, which lead to substantially different selection of optimal sampling points, there are other differences between our work and (26). In particular, only numerical simulations were provided in (26), while we have also provided phantom and in vivo data to validate the precision gains of our point selection approach directly. Furthermore, we have also evaluated our selection strategy using a Bayesian CRB to accommodate a wide range of T_1 values of interest.

From both the theoretical and experimental perspectives, it is observed that the standard deviation of the estimator increases with increasing T_1 values, which is reflected in the choice of sampling points when a range of T_1 values are of interest. This is also reflected in the corresponding gains of the optimized point selection versus uniformly distributed point selection, with more substantial gains in the precision for higher T_1 values. Conversely, for shorter T_1 values, both the percentage and the absolute reduction in the standard deviation of the estimator were smaller when using the optimized point selection.

Optimization of the sampling points for one range of T_1 values does not guarantee a systematic improvement in precision across all T_1 values. This is consistent with previous studies in parameter mapping (25,26). This was further validated in phantom imaging, where the sampling points optimized for T_1 values from 950 to 1250 ms, did worse than uniform distribution in terms of precision for low T_1 values < 400 ms. Optimizing for a larger spread of values also requires careful consideration as to how the final variance should be weighted. For instance, equal weights for T_1 values of interest of 450 and 1250 ms yielded the same distribution as just optimizing for 1250 ms as the variance of the estimator is dominated by the larger T_1 value. However, by switching to weights scaling as $1/T_1^2$, thus, characterizing the weighted error as a percentage of the T_1 values, an optimal point selection that systematically outperformed the uniform sampling point distribution in numerical and phantom experiments was obtained.

The point selection procedure is not dependent on the SNR of the fully-recovered image, a or saturation/ inversion efficiency, b, as shown in Appendix A. However, the standard deviation of the T_1 estimator scales inverse linearly with a, and with b. Thus, the percentage gain in precision between optimal point selection and uniform point distribution is constant for varying SNR, but the numerical difference becomes small for high values of a. Hence, for high-SNR sequences, the gains from the proposed point selection may not be noticeable. As b also inverse linearly scales the standard deviation, the dynamic range acquired by doing an inversion recovery T_1 mapping doubles the precision with respect to SR, assuming accurate flip angles and ignoring effects of imaging pulses until the acquisition of central k-space.

The effects of increasing K versus improving the SNR are different based on the constraints on what points on the curve can be sampled with a given sequence. In

a practical myocardial T_1 mapping scenario, where an infinite SR time can be imaged only once, going from K=5 to 15 lead to an improvement in the standard deviation of the T_1 estimator that was worse than the corresponding improvement achieved by increasing the SNR by $\sqrt{3}$. However, in an ideal scenario, where any point can be sampled multiple times (in Appendix B), going from K=5 to 10 was better for improving the precision than increasing the SNR by $\sqrt{2}$. Thus, based on the constraints on the sampling points and the sequence, the designer can perform the numerical simulations based on the theory and conclude whether to improve the SNR or increase K for improved precision in estimation.

The general estimation theoretic framework for evaluation of variance lower bounds can be used for a wide range of sequences and applications (22–26). As the differences predicted by the theory for different sampling point selection schemes match experimental values well, the bounds themselves may be useful in comparing multiple sequences in terms of their precision theoretically in a quantitative manner. Alternatively this can also be studied with Monte-Carlo simulations (31), although this does not lead to closed-form expressions, and thus, does not allow for a computationally tractable method of optimizing point selection.

In this study, we have arbitrarily chosen a SR approach utilized in myocardial T_1 mapping for validation. This sequence has certain limitations, most notably the need to apply the saturation pulse and imaging during the same R-R interval. This limits the maximum possible saturation time to T_{max} , which decreases as the heart rate increases. Thus, for higher heart rates (or larger acquisition windows), the optimal selection at T_{max} will shift to smaller values, reducing the precision of the sequence for the same number of sampling points. Furthermore, the results in Appendix B indicate that the precision can be improved if T_{max} is higher than the T_1 value of interest. Thus, precision may be improved by applying the preparation pulse(s) in one R-R interval and imaging in the consequent one, which was explored in other works (32,33).

The accuracy of the T_1 estimation using SASHA sequence with respect to a gold standard inversion recovery spin echo sequence using phantom measurements were not evaluated, as this has already been reported (10), and further studied elsewhere (12,34). Our focus is on the choice of sampling points and how this can reduce the variance of the estimator, that is, the precision, as evaluated by signal homogeneity in an ROI where T_1 is expected to be homogenous. We do note that the CRB theory presented here is for unbiased estimators, which is an assumption satisfied by the SASHA sequence with sufficient SNR and the 3-point T_1 fitting of Eq. [2] (10). The theory of CRB can be extended to biased estimators, provided that the bias term can be characterized in terms of the parameters that are being estimated (28). An interesting implication of the CRB for biased estimators is that a biased estimator does not necessarily have worse precision than an unbiased estimator, and this extension may be appropriate for alternative myocardial T₁ mapping sequences based on Look-Locker experiments (8,9,35).

We have based the probabilistic model in our CRB derivation on least-squares estimation, the most commonly used T_1 curve fitting procedure (8–12). This has one-toone correspondence with a Gaussian noise model in the images. However, the noise in the magnitude images is Rician. Thus, there is an underlying assumption that the noise in the magnitude images can be well-approximated by Gaussian noise, which is true with sufficient SNR (27). However, this assumption may not hold for images acquired close to the saturation pulse in the SASHA sequence. This, in turn, may translate to an apparent bias in the T_1 estimates if sufficiently many of the acquired images have low SNR. For instance, such cases may happen when the heart-rate of the subject is high, thus $T_{\rm max}$ is low, which was discussed in (10). In our case, this was observed in phantom experiments for vials with high T_1 values, when sampling points were selected for low T_1 values, but were used to image vials with high T_1 values. Thus, the unbiased nature of SASHA should be considered, when using the optimal selection procedure for one range of T_1 values, but imaging another higher range of T_1 values.

Our study has several limitations. We have provided a general theory for evaluating the variance of estimators in myocardial parameter mapping, but we have only evaluated this approach in saturation-recovery-based myocardial T_1 mapping, the choice of which was arbitrary. Furthermore, studies for confirming the theoretical results for other parameter mapping techniques are warranted. The in vivo data presented was acquired using breath-hold, but respiratory drift may be present among subsequent contrast-weighted images (corresponding to different sampling points along the SR curve), which can corrupt the estimated T_1 maps and their homogeneity. In the in vivo data, inflow and off-resonance effects (36) may also increase signal inhomogeneity, affecting the final reported mean and precision of the T_1 values.

CONCLUSIONS

We have used an estimation theoretic framework to evaluating the precision of a T_1 mapping technique in the presence of noise, and to choose the location and number of sampling points on the parametric magnetization curve to achieve a given level of precision. The point selection can be optimized for different values of interest of a given parameter to achieve higher levels of precision in estimation without increasing the scan time.

APPENDIX A

Eq. 8 states

$$J(a, b, T_1, \{\mathbf{x}_k\}) = \left[\mathbf{I}^{-1}\right]_{3,3}$$

= $\frac{\mathbf{I}_{11}\mathbf{I}_{22} - \mathbf{I}_{12}^2}{\mathbf{I}_{11}(\mathbf{I}_{22}\mathbf{I}_{33} - \mathbf{I}_{23}^2) - \mathbf{I}_{12}(\mathbf{I}_{33}\mathbf{I}_{12} - \mathbf{I}_{23}\mathbf{I}_{13}) + \mathbf{I}_{13}(\mathbf{I}_{12}\mathbf{I}_{23} - \mathbf{I}_{22}\mathbf{I}_{13})}.$
[A1]

We let

$$\gamma_k = e^{\frac{-x_k}{T_1}} \text{ and } \beta_k = \frac{x_k}{T_1^2}.$$
 [A2]

Thus, the numerator of [A1] becomes

$$\begin{aligned} \mathbf{I}_{11}\mathbf{I}_{22} - \mathbf{I}_{12}^{2} &= \sum_{k=1}^{K} (1 - b\gamma_{k})^{2} \sum_{l=1}^{K} (-a\gamma_{l})^{2} \\ -\sum_{k=1}^{K} (-a\gamma_{k})(1 - b\gamma_{k}) \sum_{l=1}^{K} (-a\gamma_{l})(1 - b\gamma_{l}) \\ &= a^{2} \sum_{k=1}^{K} (1 - b\gamma_{k}) \sum_{l=1}^{K} \gamma_{l}(\gamma_{l} - \gamma_{k}) \\ &= a^{2} \sum_{k=1}^{K} \sum_{l=1}^{K} \gamma_{l}(\gamma_{l} - \gamma_{k}) - a^{2} b \left(\sum_{k=1}^{K} \sum_{l=1}^{K} \gamma_{k}\gamma_{l}^{2} - \sum_{k=1}^{K} \sum_{l=1}^{K} \gamma_{k}^{2}\gamma_{l} \right) \\ &= a^{2} \sum_{k=1}^{K} \sum_{l=1}^{K} \gamma_{l}(\gamma_{l} - \gamma_{k})$$

$$= a^{2} \sum_{k=1}^{K} \sum_{l=1}^{K} \gamma_{l}(\gamma_{l} - \gamma_{k})$$

$$(A3)$$

as $\sum_{k=1}^{K} \sum_{l=1}^{K} \gamma_k \gamma_l^2 = \sum_{k=1}^{K} \sum_{l=1}^{K} \gamma_k^2 \gamma_l$ with a simple change of indices. Similarly, the denominator of A1 is given by

$$\begin{split} \mathbf{I}_{11} \left(\mathbf{I}_{22} \mathbf{I}_{33} - \mathbf{I}_{23}^{2} \right) &- \mathbf{I}_{12} (\mathbf{I}_{33} \mathbf{I}_{12} - \mathbf{I}_{23} \mathbf{I}_{13}) + \mathbf{I}_{13} (\mathbf{I}_{12} \mathbf{I}_{23} - \mathbf{I}_{22} \mathbf{I}_{13}) \\ &= a^{4} b^{2} \sum_{j=1}^{K} \left(1 - b \gamma_{j} \right) \sum_{k=1}^{K} \sum_{l=1}^{K} \left\{ \gamma_{k}^{2} \gamma_{l}^{2} \beta_{l} (\beta_{l} - \beta_{k}) + \gamma_{j} \gamma_{k} \gamma_{l} \beta_{j} \beta_{l} (\gamma_{l} - \gamma_{k}) \right\} \\ &+ \gamma_{j} \gamma_{k} \gamma_{l} \beta_{j} \beta_{l} (\gamma_{l} - \gamma_{k}) \right\} \\ &= a^{4} b^{2} \sum_{j=1}^{K} \sum_{k=1}^{K} \sum_{l=1}^{K} \left\{ \gamma_{k}^{2} \gamma_{l}^{2} \beta_{l} (\beta_{l} - \beta_{k}) + \gamma_{j} \gamma_{k} \gamma_{l} \beta_{l} (\gamma_{k} \beta_{k} - \gamma_{l} \beta_{l}) + \gamma_{j} \gamma_{k} \gamma_{l} \beta_{j} \beta_{l} (\gamma_{l} - \gamma_{k}) \right\} \\ &- a^{4} b^{3} \sum_{j=1}^{K} \sum_{k=1}^{K} \sum_{l=1}^{K} \left\{ \gamma_{k}^{2} \gamma_{l}^{2} \gamma_{l} \beta_{l} (\beta_{l} - \beta_{k}) + \gamma_{j} \gamma_{k} \gamma_{l} \beta_{l} (\gamma_{l} - \gamma_{k}) + \gamma_{j}^{2} \gamma_{k} \gamma_{l} \beta_{l} (\gamma_{l} - \gamma_{k}) \right\} \\ &= a^{4} b^{2} \sum_{j=1}^{K} \sum_{k=1}^{K} \sum_{l=1}^{K} \left\{ \gamma_{k}^{2} \gamma_{l}^{2} \beta_{l} (\beta_{l} - \beta_{k}) + \gamma_{j} \gamma_{k} \gamma_{l} \beta_{l} (\gamma_{k} \beta_{k} - \gamma_{l} \beta_{l}) + \gamma_{j} \gamma_{k} \gamma_{l} \beta_{j} \beta_{l} (\gamma_{l} - \gamma_{k}) \right\} \end{split}$$

as $\sum_{j=1}^{K} \sum_{k=1}^{K} \sum_{l=1}^{K} \gamma_k^2 \gamma_l^2 \gamma_j \beta_l^2 = \sum_{j=1}^{K} \sum_{k=1}^{K} \sum_{l=1}^{K} \gamma_j^2 \gamma_k \gamma_l^2 \beta_l^2,$ $\sum_{j=1}^{K} \sum_{k=1}^{K} \sum_{l=1}^{K} \gamma_k^2 \gamma_l^2 \gamma_j \beta_l \beta_k = \sum_{j=1}^{K} \sum_{k=1}^{K} \sum_{l=1}^{K} \gamma_j^2 \gamma_k \gamma_l^2 \beta_j \beta_l,$ and $\sum_{j=1}^{K} \sum_{k=1}^{K} \sum_{l=1}^{K} \gamma_j^2 \gamma_k^2 \gamma_l \beta_l \beta_k = \sum_{j=1}^{K} \sum_{k=1}^{K} \sum_{l=1}^{K} \gamma_j^2 \gamma_k^2 \gamma_l \beta_j \beta_l$ after change of indices.

From these expressions, we have

$$\begin{split} J(a, b, T_1, \{x_k\}) \\ = & \frac{1}{a^2 b^2} \frac{\sum\limits_{k=1}^{K} \sum\limits_{l=1}^{K} \gamma_l(\gamma_l - \gamma_k)}{\sum\limits_{j=1}^{K} \sum\limits_{k=1}^{K} \sum\limits_{l=1}^{K} \left\{ \begin{array}{c} \gamma_k^2 \gamma_l^2 \beta_l(\beta_l - \beta_k) + \gamma_j \gamma_k \gamma_l \beta_l(\gamma_k \beta_k - \gamma_l \beta_l) \\ + \gamma_j \gamma_k \gamma_l \beta_j \beta_l(\gamma_l - \gamma_k) \end{array} \right\}}. \end{split}$$

As γ_k and β_k are functions of $\{x_k\}$ and T_1 only, this shows $J(a, b, T_1, \{x_k\})$ is inversely proportional to a^2b^2 . In particular,

$$J(a,b,T_1,\{x_k\}) = \frac{1}{a^2b^2}J(1,1,T_1,\{x_k\}).$$
 [A6]

APPENDIX B

Two additional numerical simulations of theoretical interest were performed. First, for the T_1 value of 1250 ms with K=11, T_{max} was set to 1300 ms, which is higher than the trigger delay time, and the optimization procedure was repeated. Second, to compare the effects of increasing SNR versus increasing K under the assumption that any sampling point can be selected multiple times, we simulated the sequence with K=5 and K=10 for T_1 values of interest varying from 950 to 1250 ms, allowing for multiple sampling of all points, including the point at infinity.

For the set of experiments, where $T_{\rm max} = 1300$ ms, the proposed selection of points yielded (140, 140, 140, 1060, 1060, 1060, 1060, 1060, 1060, 1060, 0), whereas the uniform distribution resulted in (140, 269, 398, 527, 656, 785, 914, 1043, 1172, 1300, ∞). These lead to CRB_{prec} of 77.2 versus 89.1 ms, respectively (13.4% reduction). The distribution of points using the proposed method shows the trend of the optimization procedure to select 1 point at ∞ , 3 points at $T_{\rm min}$, and the remaining 7 points at 1060 ms.

For the second set of experiments, where any sampling point can be selected multiple times, the optimization for the sequence with K=5 yielded (140, 1265, 1265, ∞ , ∞) for a CRB_{prec} of 86.7 ms, whereas K=10 yielded (140, 140, 1220, 1220, 1220, 1220, ∞ , ∞ , ∞) for a CRB_{prec} of 60.1 ms. We note that the point selection still favors choosing points in a concentrated region, favoring multiplicities of the points with higher signal content. It is also worth noting that going from 5 to 10 points yields an improvement of 31% in the variance, which is better than acquiring the K=5 points twice for double the SNR (29%), although the difference is not large.

REFERENCES

 Bottomley PA, Hardy CJ, Argersinger RE, Allen-Moore G. A review of 1H nuclear magnetic resonance relaxation in pathology: are T1 and T2 diagnostic? Med Phys 1987;14:1–37.

- Warntjes JB, Leinhard OD, West J, Lundberg P. Rapid magnetic resonance quantification on the brain: optimization for clinical usage. Magn Reson Med 2008;60:320–329.
- Alexander AL, Hurley SA, Samsonov AA, Adluru N, Hosseinbor AP, Mossahebi P, Tromp do PM, Zakszewski E, Field AS. Characterization of cerebral white matter properties using quantitative magnetic resonance imaging stains. Brain Connect 2011;1:423–446.
- Smith SR, Roberts N, Percy DF, Edwards RH. Detection of bone marrow abnormalities in patients with Hodgkin's disease by T1 mapping of MR images of lumbar vertebral bone marrow. Br J Cancer 1992;65: 246–251.
- Hattingen E, Jurcoane A, Daneshvar K, Pilatus U, Mittelbronn M, Steinbach JP, Bahr O. Quantitative T2 mapping of recurrent glioblastoma under bevacizumab improves monitoring for non-enhancing tumor progression and predicts overall survival. Neuro Oncol 2013; 15:1395–1404.
- Mewton N, Liu CY, Croisille P, Bluemke D, Lima JA. Assessment of myocardial fibrosis with cardiovascular magnetic resonance. J Am Coll Cardiol 2011;57:891–903.
- Marwick TH, Schwaiger M. The future of cardiovascular imaging in the diagnosis and management of heart failure, part 1: tasks and tools. Circ Cardiovasc Imaging 2008;1:58–69.
- Messroghli DR, Radjenovic A, Kozerke S, Higgins DM, Sivananthan MU, Ridgway JP. Modified Look-Locker inversion recovery (MOLLI) for high-resolution T1 mapping of the heart. Magn Reson Med 2004; 52:141–146.
- Piechnik SK, Ferreira VM, Dall'Armellina E, Cochlin LE, Greiser A, Neubauer S, Robson MD. Shortened Modified Look-Locker Inversion recovery (ShMOLLI) for clinical myocardial T1-mapping at 1.5 and 3 T within a 9 heartbeat breathhold. J Cardiovasc Magn Reson 2010;12: 69.
- Chow K, Flewitt JA, Green JD, Pagano JJ, Friedrich MG, Thompson RB. Saturation recovery single-shot acquisition (SASHA) for myocardial T mapping. Magn Reson Med 2013. doi: 10.1002/mrm.24878.
- Weingartner S, Akcakaya M, Basha T, Kissinger KV, Goddu B, Berg S, Manning WJ, Nezafat R. Combined saturation/inversion recovery sequences for improved evaluation of scar and diffuse fibrosis in patients with arrhythmia or heart rate variability. Magn Reson Med 2014;71:1024–1034.
- Kellman P, Hansen MS. T1-mapping in the heart: accuracy and precision. J Cardiovasc Magn Reson 2014;16:2.
- Schulz-Menger J, Friedrich MG. Magnetic resonance imaging in patients with cardiomyopathies: when and why. Herz 2000;25:384– 391.
- 14. Arheden H, Saeed M, Higgins CB, Gao DW, Bremerich J, Wyttenbach R, Dae MW, Wendland MF. Measurement of the distribution volume of gadopentetate dimeglumine at echo-planar MR imaging to quantify myocardial infarction: comparison with 99mTc-DTPA autoradiography in rats. Radiology 1999;211:698–708.
- Deoni SC, Rutt BK, Peters TM. Rapid combined T1 and T2 mapping using gradient recalled acquisition in the steady state. Magn Reson Med 2003;49:515–526.
- 16. Liney GP, Knowles AJ, Manton DJ, Turnbull LW, Blackband SJ, Horsman A. Comparison of conventional single echo and multi-echo sequences with a fast spin-echo sequence for quantitative T2 mapping: application to the prostate. J Magn Reson Imaging 1996;6:603– 607.
- Look DC, Locker DR. Time saving in measurement of NMR and EPR relaxation times. Rev Sci Instrum 1970;41:250–251.
- Steinhoff S, Zaitsev M, Zilles K, Shah NJ. Fast T(1) mapping with volume coverage. Magn Reson Med 2001;46:131–140.
- Clare S, Jezzard P. Rapid T(1) mapping using multislice echo planar imaging. Magn Reson Med 2001;45:630–634.
- 20. Kellman P, Arai AE, Xue H. T1 and extracellular volume mapping in the heart: estimation of error maps and the influence of noise on precision. J Cardiovasc Magn Reson 2013;15:56.
- Kay S. Fundamentals of statistical signal processing, Vol. I: Estimation theory. Englewood Cliffs, NJ: Prentice Hall; 1993.
- Brihuega-Moreno O, Heese FP, Hall LD. Optimization of diffusion measurements using Cramer-Rao lower bound theory and its application to articular cartilage. Magn Reson Med 2003;50:1069– 1076.
- Poot DH, den Dekker AJ, Achten E, Verhoye M, Sijbers J. Optimal experimental design for diffusion kurtosis imaging. IEEE Trans Med Imaging 2010;29:819–829.

- 24. Sezginer A, Kleinberg RL, Fukuhara M, Latour LL. Very rapid simultaneous measurement of nuclear magnetic resonance spin-lattice relaxation time and spin-spin relaxation time. J Magn Reson 1991;92: 504–527.
- 25. Jones JA, Hodgkinson P, Barker AL, Hore PJ. Optimal sampling strategies for the measurement of spin–spin relaxation times. J Magn Reson Ser B 1996;113:25–34.
- Zhang Y, Yeung HN, O'Donnell M, Carson PL. Determination of sample time for T1 measurement. J Magn Reson Imaging 1998;8:675–681.
- Gudbjartsson H, Patz S. The Rician distribution of noisy MRI data. Magn Reson Med 1995;34:910–914.
- Gill RD, Levit BY. Applications of the van Trees inequality: a Bayesian Cramér-Rao bound. Bernoulli 1995;1:59–79.
- 29. Kraft KA, Fatouros PP, Clarke GD, Kishore PR. An MRI phantom material for quantitative relaxometry. Magn Reson Med 1987;5:555–562.
- 30. Xue H, Greiser A, Zuehlsdorff S, Jolly MP, Guehring J, Arai AE, Kellman P. Phase-sensitive inversion recovery for myocardial T1 mapping with motion correction and parametric fitting. Magn Reson Med 2013;69:1408–1420.
- Kellman P, Xue H, Hansen MS. Comparison of myocardial T1mapping protocols: accuracy and precision. In Proceedings of the

21st Annual Meeting of ISMRM, Salt Lake City, Utah, USA, 2013. p. 1394.

- 32. Weingartner S, Akçakaya M, Kissinger KV, Manning WJ, Nezafat R. Heart-rate independent non-contrast myocardial T1 mapping. In Proceedings of the 21st Annual Meeting of ISMRM, Salt Lake City, Utah, USA, 2013. p.1358.
- Slavin GS, Stanisby JA. True T1 mapping with SMART1Map: a comparison with MOLLI. In Proceedings of the 21st Scientific Meeting of ISMRM, Salt Lake City, May 2013. p.1416.
- 34. Roujol S, Weingartner S, Foppa M, et al. Accuracy and reproducibility of four T1 mapping sequences: a head-to-head comparison of MOLLI, ShMOLLI, SASHA, and SAPPHIRE. J Cardiovasc Magn Reson 2014;16(Suppl 1):O26.
- 35. Robson MD, Piechnik SK, Tunnicliffe EM, Neubauer S. T measurements in the human myocardium: the effects of magnetization transfer on the SASHA and MOLLI sequences. Magn Reson Med 2013;70: 664–670.
- Kellman P, Herzka DA, Arai AE, Hansen MS. Influence of Offresonance in myocardial T1-mapping using SSFP based MOLLI method. J Cardiovasc Magn Reson 2013;15:63.